


## LABORATORY MOISTURE METER

**UM2000-LTE**

TO DETERMINE THE MOISTURE CONTENT OF ANY KIND  
OF POWDER AND/OR GRANULAR MATERIAL




The UM2000-LTE has been designed to determine moisture content in a very simple and reliable manner. The material is dried by a round halogen quartz lamp which applies a constant heat to the sample, at one or more temperatures that have been programmed accordingly by the operator.

The method is not influenced by the effects produced by colour, density, chemical properties or absorption which can render the measurements obtained by other methods unreliable. A sample of suitable weight is placed on the scale pan inside the drying chamber.

The test parameters are programmed from the keyboard beforehand by the operator and the test starts when the relative key is pressed or when the lid is closed.

The unit has a large LED display to facilitate the reading of the data and a luminous bar to display the weight of the sample in proportion to the full scale and to monitor the weighing process.

**BEST IN CLASS FOR:****WOOD BASED PANELS:**

PB/SPB

OSB/LSB/FOSB

MDF/HDF

**MAIN FEATURES**

- RS-232 bi-directional data interface and USB for handling and storing test results and setup parameters
- Various drying modes which are easy to programme and retrieve
- Internal temperature control over a range of 50 ÷ 160 °C
- Specially designed, user-friendly software
- Calibration certified with primary gauging samples.

**ADVANTAGES**

- Elevated measuring accuracy
- Test conducted rapidly
- Good measuring repeatability
- No maintenance required.

```


MOISTURE TEST
FILE NAME:      Surface
DATE:          24/02/2011
TIME:          18:33:25
SER NO:        00AE809387
TEST NO:       1
USER NO:       000000
Result:        %ATRO MOISTURE
Heating:       Single
Interval:      10 c
Stop:          STABLE
               0.002 g
               60 sec
Start:         MANUAL
INIT MASS:     4.823 g

```

MODE	TEMP	TIME	RESULT
Single	41C	00:10	0.17 %AM
Single	62C	00:20	0.90 %AM
Single	82C	00:30	2.03 %AM
Single	99C	00:40	3.32 %AM
Single	111C	00:50	4.42 %AM
Single	111C	01:00	5.10 %AM
Single	111C	01:10	5.47 %AM
Single	110C	01:20	5.77 %AM
Single	110C	01:30	5.98 %AM
Single	111C	01:40	6.16 %AM
Single	110C	01:50	6.30 %AM
Single	110C	02:00	6.44 %AM
Single	110C	02:10	6.56 %AM
Single	111C	02:20	6.66 %AM
Single	110C	02:30	6.73 %AM
Single	111C	02:40	6.80 %AM
Single	111C	02:50	6.89 %AM
Single	111C	03:00	6.94 %AM
Single	111C	03:10	6.99 %AM
Single	111C	03:20	7.04 %AM
Single	111C	03:30	7.06 %AM
Single	111C	03:40	7.08 %AM
Single	111C	03:50	7.11 %AM
Single	111C	04:00	7.13 %AM
Single	111C	04:10	7.15 %AM
Single	111C	04:20	7.18 %AM
Single	111C	04:30	7.20 %AM
Single	111C	04:40	7.20 %AM
Single	111C	04:50	7.20 %AM
Single	111C	05:00	7.23 %AM
Single	111C	05:10	7.23 %AM
Single	111C	05:20	7.25 %AM
Single	111C	05:30	7.25 %AM
Single	111C	05:40	7.27 %AM
Single	111C	05:50	7.27 %AM
Single	111C	06:00	7.27 %AM

\*\*\*\*\* AUTO STOP \*\*\*\*\*  
LAST TEMP: 111 C  
FINAL MASS: 4.496 g  
TEST TIME: 06:00  
MASS LOST: 0.327 g  
RESULT: 7.27 %AM

## WORKING PRINCIPLE



### TECHNICAL DATA

MAX capacity	50 g
Reading division (d)	1 mg / 0.01%
Moisture resolution	0.01%
Weighing repeatability	2 mg s.d.
Moisture repeatability	10 g 0.05% sample
Min recommended weight	2 g
Environmental temperature	0 ÷ +40 °C

### AUTOMATIC CALCULATION PROGRAMMES

% of moisture in relation to the initial weight	Initial mass - dry mass / initial mass
% of solids	Dry mass / initial mass
ATRO % M	Initial mass - dry mass / dry mass
ATRO % S	Percentage of solids/dry base

### CRITERIA FOR STOPPING THE DEVICE AUTOMATICALLY

Stop when three consecutive results are identical	Interval between one readout and the next programmable from 5 to 99 sec.
Stop when three consecutive results are identical or at end of maximum time programmed	Time and readout intervals programmable from between 5 and 99 sec.

### HEATING CRITERIA

Traditional drying	Gradual heating to the final temperature within the period of time set by the operator.
Step drying method	Heating to temperature 1 for x minutes, then to temperature 2 for x minutes, then to temperature 3 for x minutes (3 stages).
Quick pre-heat	The temperature rises to 30% beyond the temperature set and then falls to the temperature required. This procedure is useful for expediting test times for certain kinds of products.